



**TATA INSTITUTE OF FUNDAMENTAL RESEARCH**

HomiBhabha Road, Colaba, Mumbai 400 005.

*National Centre of the Government of India for Nuclear Science and Mathematics  
Deemed University*

Tel : 022-22782000 / 2890/2588/2897

Email: deepak.baghele@tifr.res.in

Fax : 022-22804610

madhuram@tifr.res.in

Website : [www.tifr.res.in](http://www.tifr.res.in)

dalvidd@tifr.res.in

REF:TIFR/PD/CA18-038/PUB

September 03, 2018

To,

Vendor/Bidder

**Subject : CORRIGENDUM No. 1 Against Tender No. TIFR/PD/CA18-038/PUB  
Dt. 13.08.2018 Tender for Thermovac Chamber (Cryogenic Environment  
Chamber)**

**Ref :- Our Tender ID No. 2018\_TIFR\_350060\_1**

**Dear Bidder,**

Please refer to subject tender published on 13.08.2018, the following corrigendum to the subject tender is being issued with some changes in original tender :

Technical Specs: 1.2m (Page no. 29):

**Supplier shall ensure clear working space of 900mm in height or best possible maximum height(excluding LN2 shrouds)**

Technical Specs: 1.2 u (Pumping port) (Page no.30):

**The chamber should have two pumping ports each of size compatible to the flange CF-F 250 (appropriate to provided turbo molecular pump diameter), mounting horizontally. Pumping port must support 50kg load of Turbomolecular pump (or appropriate shelf to be made to support the load.**

Technical Specs: 1.3 c (Temperature Stability)(Page no.31):

**Required Temperature Stability is within  $\pm 2K$ (Kelvin) or best possible.**

The last date for Bid Submission is revised to **26.09.2018 (17.30 hrs)** instead of **14.09.2018 (17.30 hrs)**

The date for Bid Opening is revised to **27.09.2018 (15.00 hrs)** instead of **17.09.2018 (15.00 hrs)**.

All other terms and conditions of subject tender shall remain unchanged. This Corrigendum No. 01 is an integral part of the subject tender and a copy of the same must be submitted along with the offer duly signed and stamped.

Thanking you.

For TATA INSTITUTE OF FUNDAMENTAL RESEARCH



DEEPAK BAGHELE  
ADMN. OFFICER 'C'